

CERTIFICATE OF ANALYSIS

85X N35 (batch A)

Certified Reference Material Information

Type: LEAD / ANTIMONY (CAST)
Form and Size: Disc 38-40mm diameter
Produced by: Universal Scientific Laboratory Pty Ltd
Certified and supplied by: MBH Analytical Ltd

Assigned Values

Percentage element by weight

Element	Sn	Bi	Ag	Se	Sb	Cu
Value ¹	0.044	0.0130	0.0011	0.0004	3.42	0.0246
Uncertainty ²	0.002	0.0006	0.0001	0.0001	0.04	0.0007

Element	Ni	S	As	Cd	Zn	Fe
Value ¹	0.00058	0.007	0.201	0.00013	0.00013	(0.0001)
Uncertainty ²	0.00006	0.001	0.002	0.00002	0.00003	-

Note: values given in parentheses are not certified - they are provided for information only.

Definitions

- ¹ The certified values are the present best estimates of the true content for each element. Each value is a panel consensus, based on the averaged results of an interlaboratory testing programme, detailed on page 3.
- ² The uncertainty values are generated from the 95% confidence interval derived from the wet analysis results, in combination with a statistical assessment of the homogeneity data, as described on page 2.

Certified by:

MBH ANALYTICAL LIMITED

on 10th October 2011


C Eyeleigh